

# Supplier's declaration of conformity

For compliance levels 1, 2 and 3 in Australia and Levels of Conformity 1, 2 and 3 in New Zealand.



**Australian Government**  
**Australian Communications and Media Authority**

R003 (Oct 06)

As required by notices under:

- section 182 of the Australian *Radiocommunications Act 1992*;
- section 407 of the Australian *Telecommunications Act 1989*; and
- section 134 of the New Zealand *Radiocommunications Act 1989*.

THIS COMPLETED FORM REMAINS WITH THE SUPPLIER  
AS PART OF THE DOCUMENTATION REQUIRED FOR THE COMPLIANCE RECORD  
DO NOT RETURN TO THE ACA OR RSM

Suppliers details	
<b>Name</b> (NAME OF MANUFACTURER OR IMPORTER)	<b>ACN, ARBN, ABN or NZCN or NZ GST number:</b>
Yokogawa Australia Pty. Ltd.	ABN 36 003 888 364
<b>Address</b> (ADDRESS OF MANUFACTURER OR IMPORTER)	<b>ACMA / MED supplier code number:</b>
PRIVATE MAIL BAG NO.2024 PO NORTH RYDE NSW POSTCODE 1670	N200
Product details	
<b>Product description—brand name, model, lot, batch or serial number</b> (IF AVAILABLE)	
Yokogawa Model: EXA402G Series CONVERTER	
General specification: GS 12J5D2-E, GS 12D08E02-01E	
Instruction Manual: IM 12J5D2-01, IM 12D08E02-01E	
General Description : EEN244-A01(Rev.3)	
Emission Report : 06TR-Y5103	
Immunity Report : 06TR-Y5103, EEN244-A02	
Applicable Standards details	
<b>Standard title, number, edition and if applicable the test report number</b>	
EN 61326-1:2006 Class A Table 2 (For use in industrial locations)	
Electrical equipment for measurement, control and laboratory use – EMC requirements	
Part 1: General requirements	
Declaration	
I hereby declare that the product mentioned above complies with the above mentioned standards and all products supplied under this Declaration will be identical to the sample identified above	<b>Signature of supplier/agent</b>
	<b>Date</b>
	15 June, 2011
Sri M Inguva	
PRINT NAME	
NATIONAL QUALITY & PROCESS IMPROVEMENT MANAGER	
POSITION IN ORGANISATION	