

# Certificate of Compliance

**Certificate:** 2300657 (LR 96902)

**Master Contract:** 172608

**Project:** 2564423

**Date Issued:** October 2, 2012

**Issued to:** Yokogawa Electric Corporation

(location M21-5E)

2-9-32 Nakacho

Musashino-shi, Tokyo 180-8750

Japan

Attention: Tooru Akai

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Constantine Somow*

**Issued by:** Constantine Somow, C. Tech.

## **PRODUCTS**

**CLASS 3631 05** - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

**CLASS 3631 85** - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Two-wire analyzer, Model FLXA21\* Series, K9698\* Series, rated 24 V dc, 4-20 mA, Type 3S/4X.

Notes:

1. The above model is Pollution Degree 2, Installation Category I, powered from SELV supply.
2. Models provided with a stainless steel enclosure are additionally rated IP66.



**Certificate:** 2300657 (LR 96902)

**Master Contract:** 172608

**Project:** 2564423

**Date Issued:** October 2, 2012

---

**APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2nd Edition) - Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Second Edition)



## *Supplement to Certificate of Compliance*

**Certificate:** 2300657

**Master Contract:** 172608

*The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
2564423	October 2, 2012	Update Report to include alternative construction.
2541604	August 22, 2012	Update Report 2300657 to include an alternative gasket for Type 3S/4X & IP66 SS enclosure.
2435658	July 4, 2011	Update Report 2300657 to include an alternative construction.
2300657	October 20, 2010	Two-wire analyzer, Model FLXA21* Series, K9698* Series (C/US).